Journal of Radiotherapy in Practice

cambridge.org/jrp

Comparative analysis of setup margin calculation in cone beam CT, by van Herk formula, using two different image registration methods

Soumya Roy[1,2], Biplab Sarkar[2,3] and Anirudh Pradhan[2]

[1]Manipal Hospital, Dwarka, New Delhi, India; [2]GLA University, Mathura, UP, India and [3]Apollo Gleneagles Hospitals, Kolkata, India

Original Article

Cite this article: Roy S, Sarkar B, and Pradhan A. (2023) Comparative analysis of setup margin calculation in cone beam CT, by van Herk formula, using two different image registration methods. *Journal of Radiotherapy in Practice*. **22**(e92), 1–8. doi: 10.1017/S1460396923000122

Received: 5 April 2022
Revised: 31 July 2022
Accepted: 17 February 2023

Key words:
clip box; cone beam computed tomography; head-neck cancer; mask; radiotherapy; random error; systematic error; volumetric modulated arc therapy

Author for correspondence:
Soumya Roy, Department of Radiation Oncology, GLA University, Mathura, UP, India.
E-mail: soumyaroy91@gmail.com

ABSTRACT

Introduction: This study aimed to quantify the difference in setup margin in cone beam computed tomography (CBCT) setup imaging, utilising the van Herk formula for two different image registration methods. Two alternative techniques of registration, bony landmark (BL) matching and soft tissue matching (ST) for head and neck cancer patients, were investigated.
Methods: This study included 30 head and neck cancer patients who received a simultaneous integrated boost of 54–60–66 Gy in 30 fractions, using volumetric modulated arc treatment. A total of 867 CBCT images were acquired during patient setup and further analysed for setup margin calculation. A region of interest was described using a clip box between the reference and CBCT image to calculate the patient's positional inaccuracy in three translational directions, X, Y and Z, where X was mediolateral, Y was the cranial-caudal, and Z was the anterior-posterior direction in the patient-based coordinate system, respectively. The shifts were captured by altering the BL and ST matching, and the setup margin was calculated using the van Herk formula ($=2{\cdot}5\Sigma + 0{\cdot}7\sigma$ where Σ was the systematic and σ was the random error).
Results: The difference between bony and ST matching in most cases was observed to be 1·4 mm in all translational directions at a 95% confidence interval and <1° in all rotational directions. The rotational error was found to be below the action level (±3°); hence, no corrections related to rotational error were made. The translational setup margin for bone and ST-based registration was X (BL) = 4·6 mm, X (ST) = 4·4 mm, Y (BL) = 6·3 mm, Y (ST) = 4·7 mm, Z (BL) = 3·0 mm, Z (ST) = 3·6mm.
Conclusion: Two distinct registration approaches for head-neck patient setup did not yield any significant difference in the setup margin calculation. A suitable approach for CBCT and reference CT registration technique was required for the setup margin calculation. Confusion in selecting the correct image registration procedure can result in incorrect treatment execution. The compatibility of the two registration approaches was established in this study. Image fusion was neutralised before the second match (ST) to avoid hysteresis. For setup verification using CBCT for the head and neck region, both bone and ST registration were compatible for setup verification.

Introduction

Volumetric modulated arc treatment (VMAT) delivered treatment faster than multiple static fields of intensity-modulated radiotherapy (IMRT) without compromising the dosimetric goal.[1] So, in several treatment regions, VMAT has become the standard of care and the technique of choice for external beam radiotherapy.[2] Head and neck cancer was one treatment region where VMAT performed better than IMRT dosimetrically and in delivery time.[3] Due to clinical requirements, planning target volume (PTV) may be very much proximal to the critical organ at risk. A high dose gradient concerning distance would be created in such a case during treatment planning for VMAT. With a high gradient in the distribution, VMAT demands a proper setup through image guidance to avoid missing any clinical or subclinical disease, leading to locoregional failure.[4] Although there were limited solutions for image guidance, like two-dimensional portal imaging, stereoscopic imaging and surface guidance, the most accurate and the standard was the linear accelerator embedded in cone beam CT.[5,6] Quite a few articles showed the high accuracy of cone beam computed tomography (CBCT) and compatibility with other imaging systems.

There were various methods available for registration of cone beam CT to planning CT.[4,7,8] The choice available to the user depended on the make and model of the equipment and the licences available for the equipment. Users must evaluate the benefits and drawbacks of each technique and establish an institution-specific protocol for patient setup image registration.

© The Author(s), 2023. Published by Cambridge University Press.

The available image registration methods using CBCT and treatment planning system (TPS)-generated images for X-ray volumetric imaging (XVI) were as follows. There were clip box registration (CBR), mask-based registration, and dual registration (DR).[7,8]

(1) CBR: In this method, an region of interest (ROI) was defined on the CT image in the form of a box drawn on any of the three orthogonal views around the ROI.
(2) Mask registration (MR): This used image registration using a particular soft tissue (ST) volume. The mask sets an irregular volume of interest around the main chosen volume. The registration was restricted to the voxels inside this volume and should include the target.
(3) DR: DR was defined as the combination of CBR & MR. The average of these errors, determined from these registrations, was attributed to the final patient setup error.

Users can use the three registration methods for patient setup verification per the requirements. In our institute, the XVI (version 4.5.1) enabled manual matching and clip box-based automatic matching. For automatic matching, the options available were bone matching or grey value-based (ST matching). Cao et al. compared image registration methodology with different ROIs in lung cancer radiotherapy.[6] It was found that the automatic image matching of onboard imaging (OBI, Varian, Palo, Alto, CA) was accurate and highly reliable for detecting offset errors.

The appropriate choice of the registration between the cone beam CT and reference image depended upon the ROI. Bone matching was considered appropriate in parts covered by a bony structure or the area where target volume was determined by a bony landmark (like the spinal cord). According to Cambel et al., ST matching was considered the gold standard.[9] This study compared the efficacy of bony matching with ST matching. Also, the image registration time for bony landmark (BL) and ST matching for the head-neck site were compared. Relevant publications mentioning the benefit and limitations of bony matching in head-neck sites compared to ST matching were not found.

The final quantitative outcome of the setup verification data was to determine the setup margin (SM). The PTV was defined by a safety margin over the clinical target volume (CTV) to be vigilant of the positional error due to random and systematic shifts of the latter. The van Hark formula was the most common technique for calculating a margin over a population of patients having multiple treatment sessions. Multiple investigators justified and validated this formulation for different treatment sites and used it in all modern clinical practices.[10,11] Therefore, determination of the SM was essential to accurately delivering treatment.

This article evaluated the characteristic change in the SM as a function of the two different image registration methods for the head and neck regions receiving treatment using the VMAT delivery technique.

Novelty: During SM calculation, much literature did not consider the type of registration used for matching. However, shift values differ based on the registration process during image matching, which may result in different PTV margins for other registration processes during portal imaging. Here we have tried to find the correlation of the CBCT matching process with setup error in head-neck cancer patients. We have not found any suitable publication for margin compared with the matching methods. It was found that less time was required for bony matching compared to ST matching, maintaining same quality which was also a new finding.

Materials and Methods

Radiotherapy treatment planning and patient selection

Simulation contouring and treatment planning

All patients were simulated in head-first supine condition with both arms by the side using a five-clamp head and neck thermoplastic mask. Using the CT room's laser (DORADO, LAP GmbH, Germany), three fiducial markers were placed on the patient's anterior and lateral places during CT scan simulation. 3mm CT slices were obtained encompassing the region 2 cm above the orbit and lower level up to carina bifurcation as a standard CT simulation approach for all head-neck patients. Despite having a poitron emission tomography (PET)/CT machine, CT was taken as primary image data for all patients. PET data were acquired for some patients based on recommendations by the clinician. CT scans were used to outline all patients' tumours. Figure 1(a) and (b) showed the typical CBCT matching window in XVI. Thirty head and neck cancer patients received a simultaneous integrated boost of 54-60-66 Gy in 30 fractions. 6MV flatten beam was used for VMAT planning. Patients received daily CBCT scans throughout therapy. When shifts above the action level were established by our hospital protocol, that is, 3 mm for translational direction, and 3 ° for rotational direction, inter-fractional images were collected. If positional error exceeded the action level, another CBCT was acquired after correcting the table position.[12] Same action protocol applied for repeated imaging. Although 900 CBCT images were taken during setup, 33 could not be retrieved, making this image CBCT 867 with 1734 registrations. Several radiation therapists registered images at separate times. Radiation therapists requested that the on-call radiation oncologist and medical physicist examine the imaging and make a determination regarding the use of shift if shift data exceeded the 3 mm, 3° action threshold during registration. Inter-operator variability cannot, therefore, be ignored in this study. All study patients' VMAT treatment plans were created using Monaco TPS (V5·11·1, IMPAC Medical System, Inc., MU). After radiation oncologist approval, the planned CT dataset and treatment coordinates were transferred to the XVI console.

CBCT image acquisition procedure

In image-guided radiotherapy treatment (IGRT), XVI system with an Elekta Synergy® (Elekta Ltd, Crawley, UK) LINAC was utilised to acquire CBCT pictures after positioning the patient using TPS shift coordinates. Before treatment, image verification system checked the patient's position. Elekta Synergy® (Elekta Ltd, Crawley, UK) had a kilovolt X-ray source arm and a flat-panel detector. The linac was equipped with EPID IviewGT Elekta, based on the aSi panel XRD 1640 AL5 (PerkinElmer Optoelectronics, Fremont, CA USA). With a pixel pitch of 400 m, the active area of the sensitive layer was 409·6x 409·6 mm^2. Elekta's scanning technology was used to obtain three-dimensional CBCT (3D-CBCT) images. Image collection parameters were 120 kV, 528 mA, clockwise or counter-clockwise gantry rotation from 180° to 180°, 1 rotation per minute gantry speed, collimator cassette M20, F1 filter and 330 frames. Using XVI software, 0·2 cm CBCT images were collected and reconstructed. XVI employed Feldkamp–Davis–Kress algorithm to match CT and CBCT.[13] XVI's user guideline suggested verifying matching precision on a phantom.[14]

CBCT image matching procedure

Monaco (V5·11·1, IMPAC Medical System, Inc., MU) TPS used 3D image series to calculate the patient's setup deviance. Feldkamp's back-projection algorithm was used to create 3D

https://doi.org/10.1017/S1460396923000122 Published online by Cambridge University Press




Figure 1. (a) and 1 (b) depicted two XVI image registration procedure for head and neck patients.

datasets from 2D image data.[13] XVI could be used for automatically or manually matching patient's images. Automatic matching was possible using either bone landmarks (BL) or ST. Bony landmark available in the PTV adjacent area was considered for bony matching. The most prevalent markers were C1–2, C5–7, occiput and mandible. Grey value-based matching (ST matching) was

https://doi.org/10.1017/S1460396923000122 Published online by Cambridge University Press




Figure 2. Translational shifts along three directions with two different registration methods by box and whisker plot.

time-consuming compared to bony matching. For ST matching, the time required for a head-neck case was 28·6 ± 7·5 s. Bony landmark-based image registration took 1s for all head-neck patients under study. As per Meyer et al., ST matching was considered superior to bone matching.[15]

Grey value (translational(T)+ rotational(R)) matching

In this work, patient treatment setup errors were analysed using ST (T + R) automatic image matching procedure. Maximum translational and rotational error tolerances were 3 mm and 3°. As the linac did not have a 6D couch, the couch under study could not correct rotational inaccuracies. Inaccuracies higher than 3° required repositioning and a second CBCT. Sarkar et al. published a rotational correction action procedure that could be further utilised for a non-6D couch system.[12]

CBR matching: The image's area of interest was boxed in sagittal, coronal, and axial views. User-specified box dimensions were available. Each image set only evaluated voxels within the target's clip box. A rigorous registration technique did not account for image margins. Chamfer matching was a powerful tool for shape-based detection in cluttered images.[16]

Patient setup errors

The patient setup error was defined as the difference between the actual and predicted position of the treated region of the body. The planned position and the image data from the TPS via the Digital Image and Communication in Medicine export process were sent to the imaging console. Image sent from TPS was considered as a baseline for the patient.

Setup errors were classified into two types: systematic errors and random errors.

Systematic error (Σ): Systematic error was considered a deviation from the planned patient position to the average position for fractionated radiation therapy. The systematic error for a population of patients was determined using the standard deviation of mean errors for each patient using two distinct registration methods.[5,17]

Random error (σ): Random error was a variation between fractions during the same patient's treatment sequence. It was measured using two different registration methods as the root mean square value of the standard deviation of errors acquired for each patient.[17,18]

Analysis of mean setup errors in three translational directions using two methods

Mean displacement vector (R)

The mean displacement vector had a length equal to the shortest distance between the point's initial and final positions. It was expressed in terms of the patient's overall setup error distance, considering three translational coordinates, where, correspondingly, dx, dy and dz were deviations in the x, y, and z directions, and r was the mean displacement.

https://doi.org/10.1017/S1460396923000122 Published online by Cambridge University Press




Figure 3 (a). Translational shifts along X direction with two different registration methods. (b). Translational shifts along Y direction with two different registration methods. (c). Translational shifts along Z direction with two different registration methods.

Margin calculation

Margin calculation was done with the van Herk formula[5]; for any translational direction as

$$\text{Margin} = 2.5\Sigma + 0.7\sigma$$

where Σ was the systematic error and σ was the random error in any particular direction for any particular site.

Statistical analysis

The statistical analysis was performed on two different registration methods using the Pearson correlation test ($P < 0{\cdot}05$) to determine the P value of the analysed data using SPSS software (SPSS V.16, IBM, IL, USA). Bland–Altman analysis was performed using Microsoft Excel 2019.

Results

The final results showed similarities and contrasts in setup errors due to the influence of two different CBCT registration techniques. Variation of shifts in different directions under two different registration is shown in Figure 2 by using a box and whisker plot. It could be inferred from the plot that the deviation was within

https://doi.org/10.1017/S1460396923000122 Published online by Cambridge University Press




Figure 4. Comparison of mean setup error in two different registration methods.




Figure 5. Comparison of mean displacement vector for bone and grey value matching.

2 mm in most instances. Bland–Altman plots were used to compare two different methodologies' efficacy. Bland–Altman plots identify systematic differences (fixed bias) or outliers. Bland–Altman plot (difference plot) compared two arrays. It could also be used to reach a new measurement technique or method with a gold standard, as even a gold standard was not error-free. The shift data obtained from two methodologies were analysed using Bland–Altman analysis for three translational directions, as shown in Figure 3(a)–(c). Figure 3(a)–(c) illustrated that the two methodologies were comparable, and the difference was insignificant.

The mean setup errors for BL and ST matching in the X direction (0·48 mm, 0·75 mm), in the Y direction (0·33 mm, 0·56 mm) and in the Z directions (0·47 mm, 0·17 mm) were observed, as shown in Figure 4. For grey value matching (ST matching), the time required for a head-neck case was 28·6 ± 7·5 s. BL took 1s for all head-neck patients with same area of interest. As shown in Figure 5, the value of the mean displacement vector was lower in BL method (3·68 mm) than in ST method (4·6 mm). There was no statistical significance of one method over another ($P > 0·05$). For BL and ST, the systematic error contribution along 3 translational directions was X direction (0·6, 0·54), Y direction (2·02,1·92) and Z direction (1·3,1·15), respectively. The random error contribution for BL and ST were the X direction (1·4, 1·28), Y direction (1·35, 1·33) and Z direction (1·41, 1·33), respectively, along with three translational directions. This was shown in Table 1. The values for total systematic error (Σ) and total random error (σ) for different translational directions with different matching methodologies are shown in Table 2.

The margin calculated based on different registration methodology was as follows; X margin (BL) = 4·6 mm, X margin (ST) = 4·4 mm, Y margin (BL) = 6·2 mm, Y margin (ST) = 4·7 mm, Z margin (BL) = 3·0 mm, Z margin (ST) = 3·6 mm.

Discussion

Accurate patient positioning was crucial for clinical success. Accurate patient placement reduced CTV-to-PTV margins. Reducing the CTV-to-PTV margin allowed target volume dose escalation, which may improve tumour control and prevent normal tissue complication rate. CBCT could be used to measure and correct patient setup errors, improving RT efficacy and

https://doi.org/10.1017/S1460396923000122 Published online by Cambridge University Press

Table 1. Calculation for systematic and random part from CBCT-acquired data for two different registration methods

		Systematic error			Random error		
Category	n	x(ML) in mm	y(CC) in mm	z (AP) in mm	x(ML) in mm	y(CC) in mm	z (AP) in mm
Bone Landmark Matching	867	0·6	2·02	1·3	1·4	1·35	1·41
Grey value Matching	867	0·54	1·92	1·15	1·28	1·33	1·33

Table 2. ∑ and σ calculation for two different registration methods

	Total standard deviation of systematic error	Total standard deviation of systematic error
Error components in van Herk's formula	σ (mm)	Σ (mm)
X Direction (Bone Matching)	1·33	1·90
X Direction (Grey value Matching)	1·44	1·19
Y Direction (Bone Matching)	1·99	1·80
Y Direction (Grey value Matching)	1·39	1·74
Z Direction (Bone Matching)	0·79	1·53
Z Direction (Grey value Matching)	1·02	1·53

accuracy. IGRT reduced patient setup errors, according to various studies.[19,20]

Because complex treatment options like VMAT and IMRT produce a more conformal and steep dose distribution, IMRT and VMAT require precise patient positioning. Patient placement was vital for improving target localisation efficacy and reducing morbidity in crucial organs. Many factors contribute to uncertainty in the placement of the PTV. The van Herk formula used CBCT image shift measurements to calculate the PTV margin.[6] In head-neck VMAT treatment, there were very few published literatures comparing BL and ST matching with respect to possible impact in PTV margin. The current study analysed translational errors in setup adjustments utilising CBCT images in head-neck cancer treatment using VMAT and two registration approaches.

Hawkins et al. tested the CBR approach on oesophageal cancer patients.[20] Their study analysed 122 EPID image pairs and 207 CBCT images.[19] According to Hawkins et al., clip box orientation and registration methods can affect displacements (translation and rotation). The clip box area could fluctuate in clinical applications based on clinical site and volume.

Guckenberger et al. studied the influence of clip box size on head and neck cancer patient setup errors using 98 CBCT pictures.[20] Guckenberger et al. recommended limiting clip box size to the ROI, including the patient's skull. In VMAT breast cases, all registration methods exhibited an insignificant difference in patient setup error.[8] According to Mohandas et al.,[8] automatic CBR matching ensured easy target position verification. The insignificant difference in bony and ST matching for head-neck patients provided similar inference as Mohandas et al. for breast patients. Goldsworthy et al. compared clip box and DR to treat oropharyngeal carcinoma.[7] The results of this study demonstrated that there was a clinical difference between using a standard CBR and DR in patients with oropharyngeal cancer. Our results showed an insignificant difference between ST and bony matching.

The lack of statistical difference does not mean there is no significant difference but that the method and data cannot demonstrate a difference statistically. Hence, an alternative approach, viz. Bland–Altman analysis, was chosen to analyse the data further. The Bland–Altman analysis results are shown in Figure 3(a)–(c). In Figure 3(a)–(c), it was clearly seen that most of the data were within the upper line of agreement (LOA) and lower LOA for all their translational direction. It implied all three dataset's mean values of two different methodologies have a similar outcome or are comparable. A similar inference was obtained from the Pearson correlation test. The result from the Bland–Altman analysis inferred the same conclusion as the other methodology. The two registration techniques revealed no statistically significant differences in the systematic error component, random error component, mean displacement vector, or mean setup error. In this study, bone matching was faster than grey value matching.

Interobserver variability during image registration, patient motion during image acquisition and severe reduction in image quality due to the presence of dental prosthesis were the limitations of our study. No intrafraction images were taken into consideration in this study.

Radiotherapy departments may have the additional benefit of faster registration time with the bone registration method compared to a ST registration method. Reduced time reclining on the treatment couch relieved a patient with a tight thermoplastic mask on the face. Reducing treatment duration could reduce couch wait times and setup errors. Bony matching sped up the process of image matching for head and neck patients without sacrificing accuracy in registration. It was an addition that may be deployed routinely to provide faster treatment while keeping the same quality of care.

Conclusion

The different registration processes provided different perspectives on image registration. Shift results ended up maintaining an action protocol for shift management. Hence, it was essential to quantitatively assess other registration processes while using them in our department. The SM difference obtained during this assessment could be incorporated while delineating the PTV for any particular site. In this study, two different registration methods for head-neck cancer patients were evaluated quantitatively. PTV margin was calculated based on van Herk formulae. Mean translational setup error and mean displacement vector for two registration methods were also compared in this study. The two registration methods, namely bone matching and ST matching, did not show any statistically significant difference in their ability to detect patient positioning errors in X, Y and Z directions for head-neck VMAT treatment delivery. For patient setup verification, the user can select either one of the two registration methods. Image registration in head-neck cases using bony matching was less time-

https://doi.org/10.1017/S1460396923000122 Published online by Cambridge University Press

consuming while maintaining the same registration quality. It was an excellent feature which could be regularly utilised to deliver a faster treatment maintaining the same quality of treatment. Dependence of registration method on the PTV margin for different sites can be calculated and compared. Hence, PTV margin calculation and efficacy of each registration method for brain, thorax and pelvic sites can be further evaluated. Thereafter, the relationship between PTV margin and registration methods can be understood in a more efficient manner.

Acknowledgements. We would like to thank Mr. Abhijt Pal, a medical physicist at AMRI Hospital Kolkata, for his assistance. Mr. Nilendu Saha, an Elekta service engineer, is also appreciated for responding to my inquiries on occasion. Dr. Arjunan Manikandan, Senior Medical Physicist at the Apolo Proton Cancer Centre, also deserves our gratitude.

Financial Support and Sponsorship. Nil.

Conflicts of Interest. There are no conflicts of interest.

References

1. Munshi A, Sarkar B, Roy S, Ganesh T, Mohanti BK. Dose fall-off patterns with volumetric modulated arc therapy and three-dimensional conformal radiotherapy including the "organ at risk" effect. Experience of linear accelerator-based frameless radiosurgery from a single institution. Cancer Radiother 2019; 23 (2): 138–146.
2. Sarkar B, Manikandan A, Nandy M, Munshi A, Sayan P, Sujatha N. Influence of Monte Carlo variance with fluence smoothing in VMAT treatment planning with Monaco TPS. Indian J Cancer 2016; 53 (1): 158–161.
3. Broggi S, Perna L, Bonsignore F, et al. Static and rotational intensity modulated techniques for head-neck cancer radiotherapy: a planning comparison. Phys Med 2014; 30 (8): 973–979.
4. Teoh M, Clark CH, Wood K, Whitaker S, Nisbet A. Volumetric modulated arc therapy: a review of current literature and clinical use in practice. Br J Radiol 2011; 84 (1007): 967–996.
5. van Herk M Errors and margins in radiotherapy. Semin Radiat Oncol 2004; 14 (1): 52–64.
6. Cao X, Liu M, Zhai F, et al. Comparative evaluation of image registration methods with different interest regions in lung cancer radiotherapy. BMC Med Imaging 2019; 19 (1): 100.
7. Goldsworthy S, Leslie-Dakers M, Higgins S, et al. A pilot study evaluating the effectiveness of dual-registration image-guided radiotherapy in patients with oropharyngeal cancer. J Med Imaging Radiat Sci 2017; 48 (4): 377–384.
8. Mohandass P, Khanna D, Kumar TM, et al. Study to compare the effect of different registration methods on patient setup uncertainties in cone-beam computed tomography during volumetric modulated arc therapy for breast cancer patients. J Med Phys 2018; 43 (4): 207–213.
9. Campbell A, Owen R, Brown E, Pryor D, Bernard A, Lehman M. Evaluating the accuracy of the XVI dual registration tool compared with manual soft tissue matching to localise tumour volumes for post-prostatectomy patients receiving radiotherapy. J Med Imaging Radiat Oncol 2015; 59 (4): 527–534.
10. Gill SK, Reddy K, Campbell N, Chen C, Pearson D. Determination of optimal PTV margin for patients receiving CBCT-guided prostate IMRT: comparative analysis based on CBCT dose calculation with four different margins. J Appl Clin Med Phys 2015; 16 (6): 252–262.
11. Dionisi F, Palazzi MF, Bracco F, et al. Set-up errors and planning target volume margins in head and neck cancer radiotherapy: a clinical study of image guidance with on-line cone-beam computed tomography. Int J Clin Oncol 2013; 18 (3): 418–427.
12. Sarkar B, Ganesh T, Munshi A, et al. Rotational positional error-corrected linear set-up margin calculation technique for lung stereotactic body radiotherapy in a dual imaging environment of 4-D cone beam CT and ExacTrac stereoscopic imaging. Radiol Med 2021; 126 (7): 979–988.
13. Schäfer D, Borgert J, Rasche V, Grass M. Motion-compensated and gated cone beam filtered back-projection for 3-D rotational X-ray angiography. IEEE Trans Med Imaging 2006; 25 (7): 898–906.
14. XVI R4.5.0 Instructions for Use: Elekta Synergy®, Elekta Axesse™, Elekta Infinity™, Versa HD™. Crawley, United Kingdom: Elekta Limited; 2013.
15. Meyer J, Wilbert J, Baier K, et al. Positioning accuracy of cone-beam computed tomography in combination with a HexaPOD robot treatment table. Int J Radiat Oncol Biol Phys 2007; 67 (4): 1220–1228.
16. Yu Q, Wei H, Yang C. Local part chamfer matching for shape-based object detection. Pattern Recognit 2017; 65: 82–96.
17. Czajkowski P, Piotrowski T. Registration methods in radiotherapy. Rep Pract Oncol Radiother 2019; 24 (1): 28–34.
18. Palombarini M, Mengoli S, Fantazzini P, Cadioli C, Degli Esposti C, Frezza GP. Analysis of inter-fraction setup errors and organ motion by daily kilovoltage cone beam computed tomography in intensity modulated radiotherapy of prostate cancer. Radiat Oncol 2012; 7: 56.
19. Hawkins MA, Aitken A, Hansen VN, McNair HA, Tait DM. Set-up errors in radiotherapy for oesophageal cancers–is electronic portal imaging or cone beam more accurate? Radiother Oncol 2011; 98 (2): 249–254.
20. Guckenberger M, Roesch J, Baier K, Sweeney RA, Flentje M. Dosimetric consequences of translational and rotational errors in frame-less image-guided radiosurgery. Radiat Oncol 2012; 7: 63.

https://doi.org/10.1017/S1460396923000122 Published online by Cambridge University Press