The Journal of Laryngology & Otology

cambridge.org/jlo

The hidden cochlear implant

O Ergun[1], O Yildirim[2], I Bozyel[2], I Kaymak[2], D Gokcen[2] and L Sennaroglu[3]

[1]Department of Otorhinolaryngology Head and Neck Surgery, Baskent University Hospital, Ankara, [2]Department of Electrical and Electronics Engineering, Faculty of Engineering, Hacettepe University, Ankara and [3]Department of Otorhinolaryngology Head and Neck Surgery, Faculty of Medicine, Hacettepe University, Ankara, Turkey

Abstract

Objective. The hidden cochlear implant concept has two data transmission methods: Bluetooth low energy and transtympanic optical data transfer systems. This study aimed to present the hidden cochlear implant and compare the test results with the existing fully implanted cochlear implant.

Method. The Bluetooth low energy module was implanted into the implant bed. For the transtympanic optical data transfer tests, a receiver was passed through the posterior tympanotomy, and the transmitter was placed in the ear canal.

Results. The Bluetooth low energy module range was 5.2–17.5 m. Transtympanic optical data transfer reached a rate of 1 Mbit/s and had 99.22 per cent accuracy. Despite various obstacles, the accuracy of the transtympanic optical data transfer was more than 99 per cent with a 250 Kbit/s rate. The average power consumption was 310 mW for the implanted Bluetooth low energy module and 41 mW for the transtympanic optical data transfer receiver.

Conclusion. Bluetooth low energy is suitable to be used transcutaneously. Transtympanic optical data transfer is an effective and promising technology. Hidden use cochlear implants aim to have the aesthetics of a fully implantable cochlear implant with higher reliability and a magnet-free design with smart device integration.

Introduction

Modern cochlear implants with continuously improving hardware, speech processors and atraumatic surgical techniques are a miraculous solution for treating severe sensorineural hearing loss. Because conventional cochlear implants have quite noticeable external hardware, users have demanded a cochlear implant system that does not stigmatise their disability. Although the fully implantable cochlear implant concept is not exactly new, it has not gained widespread use because of technical difficulties with rechargeable batteries, high-fidelity microphones, over-complicated surgical procedures and concerns about safety.[1]

A fully implantable cochlear implant needs to have energy and sound data transfer modalities, often with backups. Although the power supply or energy consumption of a fully implantable cochlear implant system is not the primary focus of this study, it could be said that the unavailability of a small, reliable rechargeable battery with enough capacity[2] and a long lifetime or an effective energy harvesting system has been problematic. Regardless of the technology, the low energy requirement is always critical.

The sound source of a fully implantable cochlear implant could be an implanted subcutaneous microphone,[3,4] a piezoelectric vibration sensor located on the ossicles,[5] tympanic membrane[6] or under the basilar membrane,[7] or could even use electrocochleography.[8] Although an external microphone that is wirelessly linked to a cochlear implant system without any external unit, as proposed in this study, cannot be classified as a fully implantable cochlear implant, a similar 'hidden operation' is still possible. The processing power of a modern smart device is more than enough to perform any digital sound data process almost instantly.

The simplicity and flexibility of using the microphone of a smart device and gathering the processed sound data through a wireless link, such as Bluetooth™, are very appealing. Furthermore, a readily available smart device connection would dramatically enhance the user experience, making software upgrades practical and remote control of the device possible. The drawback of a Bluetooth connection is the higher energy consumption of the implanted unit. Therefore, an energy-efficient way of transmitting sound data as an alternative would be necessary. For this purpose, we proposed a 'hidden external unit' with an in-the-canal hearing aid form and a transtympanic optical data transfer system. The hidden external unit would gather sound signals with its in-the-canal microphone. This would enable the resonance of the outer ear to give more natural hearing. The processed sound data would then be transmitted with a wide-angle infrared source. The hidden cochlear implant would receive those signals with its optical receiver. A tiny (sub-millimetre[9]) optical receiver would be neatly embedded in the polymeric sheath of the electrode carrier, a few milimeters proximal to the round window insertion marker. Therefore, it would be positioned in the middle ear, facing the tympanic

Main Article

Dr O Ergun takes responsibility for the integrity of the content of the paper

Presented at the 42nd Turkish National Otorhinolaryngology Head and Neck Surgery Congress, 3–7 November 2021, Kyrenia, Cyprus.

Cite this article: Ergun O, Yildirim O, Bozyel I, Kaymak I, Gokcen D, Sennaroglu L. The hidden cochlear implant. *J Laryngol Otol* 2023;**137**: 1207–1214. https://doi.org/10.1017/S0022215123000130

Accepted: 10 January 2023
First published online: 8 February 2023

Key words:
Cochlear Implants; Inner Ear; Quality Of Life

Corresponding author:
Dr O Ergun,
Department of Otorhinolaryngology – Head and Neck Surgery,
Baskent University Hospital,
Maresal Fevzi Cakmak Ave. no 45 Cankaya,
Ankara 06490, Turkey
E-mail: drergun@gmail.com

© The Author(s), 2023. Published by Cambridge University Press on behalf of J.L.O. (1984) LIMITED

membrane in line of sight of the transmitter. So, the form factor and the implantation of the hidden cochlear implant would be almost identical to conventional cochlear implants.

This study aimed to present the fresh frozen cadaver implantation results of the hidden cochlear implant proof of the concept prototype and compare it with the state-of-the-art fully implantable cochlear implant concept.

Materials and methods

The experimental setup used Bluetooth low energy technology. ESP32SoC modules (Espressif Systems, Shanghai, China) were used both as the transmitter and the receiver. The transmitter module was implanted in a sub-periosteal pocket with a retro-auricular incision where a cochlear implant body would normally be placed (Figure 1). A DP832 power supply (Rigol, Suzhou, China) was used to monitor the power consumption. Transmitted and received data were continuously recorded. The average delay between the transmitted–received signals and the power consumption was calculated at 1 m. All electronic devices and wireless connections that might interfere with the results had been turned off.

Received signal strength indicator tests were performed with NRF Connect android application (Nordic Semiconductor, Trondheim, Norway) running on a Galaxy M31 smartphone (Samsung, Suwon-si, South Korea) and X509JB laptop computer (Asus, Taipei, Taiwan). Multiple ($n = 10 \pm 2$) decibel milliwatt readings were taken at each test. The schematic representation of the Bluetooth low energy data transmission system is given in Figure 2. For the distance received signal strength indicator tests, the distance between the devices was gradually increased with a clear line of sight until a reliable connection could not be established.

Obstacle and interference received signal strength indicator tests were performed at 1 m. The signal strength from the opposite side of the cadaveric head, through 8 layers of cotton-polyester blend surgical cloth (4.45 mm in total), through a 31-mm thick living tissue (palm of a hand), and through grounded and non-grounded 70-micron thick copper-clad laminate (Pyralux® AP, DuPont, Wilmington, USA), was measured. The interference test setup, which had the head in the centre and four interfering devices with Bluetooth connections at the corners of a square with 75-cm long edges, was designed according to the literature (Figure 1).[10]




Figure 1. Image of the general layout during the interference received signal strength indicator test. The Bluetooth low energy transmitter was implanted in the implant bed. It was connected to a computer with a flat cable. Four other wireless devices operating at the 2.4 GHz band were positioned at the corners of a 75-cm square to cause interference. The receiver was at the 1 m distance approximately at the position of the camera.




Figure 2. Schematic representation of the proposed Bluetooth low energy and transtympanic optical data transmission systems for the hidden cochlear implant concept. Transtympanic data transmission was performed with digital communication. Please note that the grey boxes were not active during the experiment. LED = light-emitting diode

https://doi.org/10.1017/S0022215123000130 Published online by Cambridge University Press

The transtympanic optical data transfer system had two components: an external transmitter and an implanted receiver. The transmitter had a high-speed infrared light-emitting diode (LED; VSMB10940, Vishay, USA). Digital signals were chosen to make the signal amplitude (light intensity) irrelevant during the transmission because digital signals are either '0' or '1'. The combination of infrared light and digital signals for the transtympanic data transmission was hypothesised to make the system less susceptible to misalignments, cerumen, effusion and so on. The receiver module had an off-the-shelf optical receiver (BPW34fs, Osram, Munich, Germany) with the dimensions of $5.4 \times 4.3 \times 3.2$ mm and a comparator (LM311, Texas Instruments, Dallas, USA). The schematic representation of the system is given in Figure 2.

In the experimental setup, the infrared transmitter (940 nm wavelength) LED was wired to a waterproof cable, embedded in silicone, passed through a modified earplug and facing the tympanic membrane. Similarly, the infrared optical receiver was also wired to a waterproof cable and embedded in silicone. The optical receiver was passed through the posterior tympanotomy to be located on the promontorium, facing the tympanic membrane. A conventional, post-auricular, trans-mastoidal cochlear implantation technique and standard surgical pieces of equipment were used. Anatomical landmarks were meticulously preserved (Figure 3). Other components of the transmitter and receiver were left resting on the bench. Digital data transmission, power consumption and transtympanic optical data transfer with obstacles were tested. Transtympanic optical data transfer tests were conducted with data transfer rates of 1 Mbit/s and 250 Kbit/s.

For the transtympanic optical data transfer test, both the LED driver of the transmitter and the comparator of the receiver were connected to an oscilloscope to monitor the system performance. Both the transmitter and the receiver modules were powered by a DP832 power supply. Voltage, current and the power consumption of the modules were recorded for 1 minute during the transtympanic optical data transfer test to calculate average power consumption. The transtympanic optical data transfer tests with obstacles were designed to test the system performance in suboptimal conditions, simulating real-life use. For this purpose, water filling or some disc-shaped (diameter = 6 mm) obstacles on the intact tympanic membrane were used to simulate cerumen, effusion and foreign materials. A 0.4-mm polydimethylsiloxane, 0.1-mm polyvinyl alcohol, 0.4-mm black cardboard, 1.5-mm polylactic acid and a 0.4-mm thermoplastic polyurethane disc were tested as obstacles.

Results

Fresh frozen cadaver had 7-mm-thick soft tissue covering the implantation pocket. The Bluetooth low energy system used a 91 mA current on average during both active transmitting and receiving. The current consumption decreased as low as 5 µA during the sleep cycles. The average power consumption was 310 mW. The power consumption was only 16.5 µW during the sleep cycles. The average delay between sent and recovered data packs was 1.39 ms (± 0.064).

The received signal strength at zero distance in the open air was −26 dBm. This value dropped to −47 dBm when the transmitter was implanted subcutaneously. The signal strength decreased exponentially with increasing distance (Figure 4). A reliable connection could be established up to 17.5 m.




Figure 3. (a) The transtympanic optical data transmission test layout. The optical transmitter was wired to a waterproof cable (blue), embedded in silicone, passed through a modified ear plug (yellow) and facing the tympanic membrane. Similarly, the optical receiver was also wired to a waterproof cable (black) and embedded in silicone. The optical receiver was passed through a retro-auricular incision, mastoidectomy cavity and posterior tympanotomy, similar to the conventional cochlear implantation technique and located on the promontorium, facing the tympanic membrane. (b) Anatomical landmarks such as ear canal, facial nerve (black arrowhead), chorda tympani nerve, tympanic membrane (white arrowhead), incudal buttress (asterisk), ossicles, round window (rw) niche and promontorium (pr) were meticulously preserved. (c) The optical receiver in the middle ear through the posterior tympanotomy.

https://doi.org/10.1017/S0022215123000130 Published online by Cambridge University Press




Figure 4. Graph showing change of signal strength with distance. The Bluetooth low energy module transmitter was implanted in the implant bed, and the distance of the receiver was gradually increased with a clear line of sight. A connection could be established up to 17.5 m. RSSI = received signal strength indicator

As expected, the signal strength decreased with various obstacles that were put between the implanted transmitter and the receiver except for the copper-clad laminate, which increased the signal strength by 1.52 dBm. Once the same copper-clad laminate was grounded, it caused a 4.25 dBm weakening. The presence of the head between the transmitter and the receiver modules (head shadow effect) caused a remarkable 10.71 dBm decrease in the signal strength. The palm of the author's hand, as an obstacle, caused a 9.85 dBm drop in the signal strength. Interference of other wireless devices operating at 2.4 GHz affected the signal strength considerably, decreasing the signal strength by 8.3 dBm. Eight layers of cloth representing thick headwear caused only a 4.3 dBm decrease in the signal strength. The results of the obstacle and interference received signal strength indicator tests are summarised in Table 1. The expected Bluetooth low energy ranges with the tested obstacle and interference variables are summarised in Table 1.

Transtympanic optical data transfer tests showed impressive data transfer rates exceeding 1Mbit/s with 99.22 per cent overall bit accuracy. The average delay between the sent and received data packs was 800 ns (Figure 5). During transtympanic optical data transfer tests, the average power consumption was 35 mW for the transmitter module and 41 mW for the receiving module. Because low energy consumption was not the aim, the infrared LED of the transmitter operated at its peak power (29.7 mW) throughout the experiment. The majority of the receiver module's energy consumption was because of the operational amplifier, which used 24.75 mW. The infrared receiver used only 231 µW.

During the transtympanic optical data transfer test with obstacles, a 0.4-mm-thick black cardboard obstacle gave the best result (99.65 per cent; Figure 6), which was as good as the transtympanic optical data transfer test result without any obstacle. Although small amounts of water were present because of the melting of the cadaver throughout the experiment, filling the ear canal with water decreased the correlation ratio by more than 12 per cent (correlation ratio, 87.58 per cent). The lowest correlation ratio was with the 0.4-mm-thick thermoplastic polyurethane disc (correlation ratio, 61.02 per cent). Lower transtympanic optical data transfer rates dramatically increased accuracy, and more than 99 per cent accuracy was reached with all of the obstacles with a 250 Kbit/s transtympanic optical data transfer rate. The thicknesses, the absorbance co-efficients of the obstacles and accuracy figures with 1 Mbit/s and 250 Kbit/s rates are summarised in Table 2.

Table 1. Measured signal strengths at 1 m with various obstacles or interference

Obstacle or interference	RSSI (dBm)	Maximum Bluetooth low energy range (metres)
Line of sight	−74.216	17.5
Covered with a conductor	−72.71429	20
Under 8 layers of fabric	−77.71429	11.6
Covered with a grounded conductor	−78.35714	10.93
Interference	−82.35714	8.75
Under 31-mm-thick living tissue	−83.85714	7.29
Opposite side of the head	−84.77629	5.18

The maximum estimated ranges were calculated according to the distance received signal strength indicator (RSSI) and obstacle or interference RSSI test results

Discussion

In a fully implantable cochlear implant, implanted electronics, such as microchips, batteries, microphones and communication modules, inevitably increase the risk of a device failure. Therefore, many fully implantable cochlear implant concepts, like Envoy Medical Acclaim® often implement or hypothesise a connector[11] to enable hardware upgradeability and overcome reliability issues.[1] Hidden use cochlear implants would also have a connector to be located in the mastoid cavity to allow easier device replacement without touching the electrode array.

https://doi.org/10.1017/S0022215123000130 Published online by Cambridge University Press




Figure 5. It could be seen that 5 bytes of (a) transmitted data were (b) received without major distortion during the transtympanic optical digital data transmission test. The average delay between the sent and received data was 800 ns. Please note that the amplitude of the square signal was always the same, representing a digital signal.




Figure 6. It could be seen that the (a) transmitted data was (b) received with some minor distortion during the transtympanic optical digital data transmission test while 0.4-mm thick black cardboard was on the tympanic membrane as an obstacle. The presence of the obstacle increased the rise time but did not affect the decoded signal. The average delay between the sent and recovered data was 1 μs.

Fully implantable cochlear implant concepts often have backup strategies to overcome reliability concerns. For example, the Med-El™ Mi2000 totally implantable cochlear implant device retains a receiving antenna and magnet for charging and as a backup, similar to a conventional cochlear implant.[12] This conservative approach enables the fully implantable cochlear implant system simply to be used as a conventional cochlear implant if needed. Since inductive charging is more forgiving to misalignments and interrupted charging is not a problem with modern rechargeable batteries, the only remaining function of a magnet in a fully implantable cochlear implant system would be holding the charger or backup external unit in place. We do not believe this insurance pays off against the benefits of a magnet-free design. In the magnet-free hidden cochlear implant concept, the function of the receiving antenna would be power transmission, and the data transmission would be taken care of by the Bluetooth low energy and transtympanic optical data transfer systems. The magnetic resonance imaging compatibility of a magnet-free cochlear implant would be revolutionary.

https://doi.org/10.1017/S0022215123000130 Published online by Cambridge University Press

Table 2. Correlation ratios of transmitted and received signals during the digital data transmission test with obstacles

Obstacle material thickness[27]	Optical absorbance at 940 nm (calculated for thickness)	1 Mbit/s correlation ratio (%)	250 Kbit/s correlation ratio (%)
Polydimethylsiloxane 0.4 mm	<0.001[22]	90.67	100
Black cardboard 0.4 mm	0.13[24]	99.65	100
Black cardboard 0.4 mm & water	0.2[21,23]	87.58	99.65
Polyvinyl acetate 0.1 mm	0.086[26]	72.08	99.50
Polylactic acid 1.5 mm	0.35[25]	69.02	99.35
Thermoplastic polyurethane 0.4 mm	0.40[27]	61.02	99.22

The optical absorbance capabilities of the obstacles were negligible at 940 nm. Absorbance co-efficients were taken from the literature. Please note that at a 250 Kbit/s transfer rate, correlation ratios were dramatically increased

The requirements of data and power transmission systems are usually conflicting. Conventional cochlear implants have to compromise one for the sake of the other. If data transmission was not required, optimising a wireless power transmission system would be much easier. An inductive charging system running at its resonance frequency with a narrow bandwidth could have a much higher efficiency[13] than conventional cochlear implants, which have around 40 per cent power transfer efficiency.[14]

Each sound sensor modality for fully implantable cochlear implant has its distinct pros and cons. A subcutaneous microphone could be a straightforward solution but may suffer from attenuation and distortion of the signal because of headwear, movements and chewing.[1,3,4] Piezoelectric vibration sensors may require delicate implantation surgery and be fragile.[5] A sensor array in contact with the tympanic membrane might have the risk of being exposed.[1,15] An electrocochleography-based sound sensor concept[8] seems interesting but may only be useful in patients with residual hearing, enough hair cell function and who are electroacoustic-stimulation candidates.

The potential benefits of a visual user interface and high-speed internet connection through a smart device cannot be emphasised enough. Therefore, we consider a wireless connection ability crucial for a modern cochlear implant. Bluetooth low energy was the preferred wireless technology for the hidden cochlear implant because of its lower energy consumption, higher range and shorter connection latency. A central bluetooth low energy device may connect to multiple peripherals at once. This enables an implanted unit to be controlled or monitored from one device and receive sound data from one of the multiple sources. According to the product datasheet, the maximum throughput of Bluetooth low energy communication between ESP32Soc modules can reach 700 Kbit/s.[16] Bluetooth low energy also uses 128-bit encryption for safety. Despite these advantages, Bluetooth low energy uses too much energy during active transmission and receiving, especially under interference. In our experiment, because the focus of this study was not power efficiency, we used commercial off-the-shelf Bluetooth low energy modules with the default signal strength (0 dBm). The average power consumption was 310 mW during active transmission or receiving, which decreased 18 200 times during the deep sleep cycles.

Within a power-efficiency oriented integrated chip, a Bluetooth low energy core with its peripherals would realistically consume 9–15 mW.[17] That would still be too high for a practical cochlear implant to last a full day. For this reason, the transtympanic optical data transfer concept was hypothesised in the hidden cochlear implant to maximise battery life.

The implanted module had a respectable 17.5 m range through the 7-mm soft tissue of the fresh frozen cadaver (Figure 4). The effects of the obstacles and interference were tested to give an idea of real-world conditions. Eight layers of a thick fabric simulating heavy headwear caused a slight decrease in the range. Interference almost halved the range to 8.75 m. Placing the head or the palm between the transmitter and the receiver created noticeable effects, but even in those cases, the range was more than 5 m, which would be enough. Conductor materials even had a positive effect unless grounded (Table 1). These results showed that the Bluetooth low energy range would not be a significant problem for an implanted cochlear implant.

The transtympanic optical data transfer concept is not new. It had been considered among the suitable technologies when the first transcutaneous multichannel cochlear implant concepts emerged. White advocated a cochlear implant concept that had a transtympanic optical link for wideband data transfer and a separate induction link for energy transfer.[18] Unfortunately, this concept has never been further developed. Modern rechargeable battery technology, Bluetooth, advanced electronics and the inclusion of a connector could open a second chapter for transtympanic optical data transfer.

Transtympanic optical data transfer could considerably decrease the battery consumption of the implanted unit by letting the Bluetooth low energy module sleep most of the time. With the transtympanic optical data transfer managing the routine sound data transfer, the Bluetooth low energy usage could be reduced to the initial transtympanic optical data transfer calibration step and data-logging with regular intervals, which may be as sparse as once an hour. This usage would not have a noticeable impact on the battery. But since Bluetooth low energy enabled two-way communication, it would still be needed for the concept.

An analogue transtympanic optical data transfer system could work in ideal conditions, where the receiver and the transmitter were face to face without any obstacles. But the real-life conditions are far from the 'ideal'. There would always be misalignments, movement and obstacles, such as cerumen or effusion. These changes would affect the received light intensity (signal amplitude), therefore distorting analogue data. In order to make the hidden cochlear implant immune to such variations, we preferred digital signal transmission and chose infrared light because of its deeper penetration.[19]

One of the main objectives of the experiment was to reach the highest possible transtympanic optical data transfer rate. A 1Mbit/second data transfer rate could be achieved in optimal conditions. That rate was almost three times the compact disc quality sound, more than most conventional cochlear

https://doi.org/10.1017/S0022215123000130 Published online by Cambridge University Press

implants[14] and comparable to Bluetooth low energy.[16] Such a high rate would be unnecessary for a hidden cochlear implant. Therefore, we decreased the transfer rate to enable longer pulse widths. At the 250 Kbit/s rate, which would be more than enough, the effect of optical scattering was reduced to a minimum, and accuracy increased dramatically (Table 2).

Conventional cochlear implant communication uses a radio frequency link, which is not digital and has low noise and interference immunity. Because exact accuracy figures have not been shared publicly (vague terms like 'high accuracy' are used),[14] it can be estimated to be on the high end of the spectrum of 87.8 to 96.0 per cent as given for a different radio frequency communication application.[20]

Considering the 99.21 per cent correlation ratio of the transtympanic optical data transfer test with the 1Mbit/s rate, our results were very satisfactory for a proof of the concept study. Transtympanic optical data transfer tests with obstacles proved that transtympanic optical data transfer could also be accomplished without a clear line of sight. Even with the thickest obstacle (1.5-mm-thick polylactic acid disc), which would be unlikely in a real-life scenario, 99 per cent accuracy could be reached.

Infrared light of 940 nm has low absorbance co-efficients with many polymeric substances.[21–27] For example, polydimethylsiloxane, a potential polymer to cover the infrared receiver, has a negligible absorption co-efficient.[22] Water has an absorption co-efficient of 0.7/m.[28] This means a water cover of 5 mm would absorb approximately 3.5 per cent of the light and transmit the rest. Therefore, we believe that optical scattering played a more significant role with the tested material thicknesses. This was supported by the increased accuracy with increased pulse widths.

- Bluetooth low energy technology is a viable option for implanted units especially if a connector is implemented to allow replacement and technology upgrades
- Transtympanic optical data transmission is an effective and applicable concept with an excellent low-power potential
- Hidden use cochlear implants are an innovative cochlear implant concept that incorporates Bluetooth low energy and transtympanic optical data transfer technologies

Since the power consumption of the implanted unit was important, it was worth noting that the infrared receiver only drew 70 µA. That low consumption could give an idea about the potential of the technology. Nevertheless, our study was not designed to achieve the lowest possible power consumption. All the electronic components used, such as the Bluetooth low energy module, operational amplifier, comparator, LED, receiver detector, and so on were off-the-shelf products. Therefore their consumption could not be decreased according to the specific needs of the hidden cochlear implant. Still, transtympanic optical data transfer consumed noticeably lower power (41 mW) than the Bluetooth low energy module (310 mW). It is known that energy consumption could be decreased up to one-thousandth of the current level with an integrated chip design.[29,30]

Tissue heating could be discussed as a potential risk of an infrared emitter. But the power efficient LED technology and the deep penetration of infrared reduces the risk of a harmful local heating. Although full power of the transmitter was used because decreasing the infrared radiation was not an aim of this study, we did not notice an increased tissue defrosting or heating. A 450 mW infrared laser, which is more than 10 times stronger than our setup, has been used as a therapeutic alternative in chronic tinnitus with no patient discomfort.[31] Furthermore, in the hidden cochlear implant concept, the implanted unit would be able to communicate with the hidden external unit through the Bluetooth low energy module and give feedback about the signal strength and any possible misalignments during initial calibration. This could decrease energy consumption by reducing any unnecessary light intensity and minimising tissue heating.

Conclusion

In the future, the potential of the hidden cochlear implant concept should be further explored with an application-specific integrated chip design and custom transmitters and receivers. Various signal modulation approaches should be tested to further improve the accuracy in suboptimal conditions while decreasing the power consumption.

Using off-the-shelf electronics was a major limitation of this study. This prevented accurate inferences on the accuracy, delay and especially power consumption. Because a custom-designed integrated chip would significantly improve these elements, our results could be accepted as benchmarks for, hopefully, much better future results. Another shortcoming of the study was implanting in non-living tissue. It could be said that blood flow and motion would disturb the results. Although precautions were taken, these should be tested further once an integrated chip prototype is developed and implanted in an animal model.

Competing interests. Some of the authors (OE, DG, LS) have a patent on a cochlear implant system with a connector (PCT/TR2020/050419). This patent has not been linked to any financial interest.

References

1 Cohen N. The totally implantable cochlear implant. *Ear Hear* 2007;**28**:100–1
2 Liang Y, Zhao C-Z, Yuan H, Chen Y, Zhang W, Huang J-Q *et al.* A review of rechargeable batteries for portable electronic devices. *InfoMat* 2019;**1**:6–32
3 Briggs RJ, Eder HC, Seligman PM, Cowan RS, Plant KL, Dalton J *et al.* Initial clinical experience with a totally implantable cochlear implant research device. *Otol Neurotol* 2008;**29**:114–9
4 Zenner HP, Leysieffer H, Maassen M, Lehner R, Lenarz T, Baumann J *et al.* Human studies of a piezoelectric transducer and a microphone for a totally implantable electronic hearing device. *Am J Otol* 2000;**21**:196–204
5 Barbara M, Filippi C, Covelli E, Volpini L, Monini S. Ten years of active middle ear implantation for sensorineural hearing loss. *Acta Otolaryngol* 2018;**138**:807–14
6 İlik B, Koyuncuoğlu A, Uluşan H, Chamanian S, Işık D, Şardan-Sukas Ö *et al.* Thin film PZT acoustic sensor for fully ımplantable cochlear ımplants. *Proc* 2017;**1**:366
7 Liu Y, Zhu Y, Liu J, Zhang Y, Liu J, Zhai J. Design of bionic cochlear basilar membrane acoustic sensor for frequency selectivity based on film triboelectric nanogenerator. *Nanoscale Research Letters* 2018;**13**:191
8 Riggs WJ, Hiss MM, Skidmore J, Varadarajan VV, Mattingly JK, Moberly AC *et al.* Utilizing electrocochleography as a microphone for fully ımplantable cochlear ımplants. *Sci Rep* 2020;**10**:3714
9 Excelitas Technologies Corp. In: https://www.excelitas.com/product/c30724eh-si-apd-500um [24 December 2022]
10 Cantón Paterna V, Calveras Augé A, Paradells Aspas J, Pérez Bullones MA. A Bluetooth low energy indoor positioning system with channel diversity, weighted trilateration and kalman filtering. *Sens* 2017;**17**:2927
11 Envoy Medical. In: https://www.envoymedical.com/acclaim-cochlear-implant [8 July 2022]
12 Med-El. In: https://www.medel.com/press-room/press-list/press-details/2020/10/19/first-surgeries-ever-in-europe-with-a-totally-implantable-cochlear-implant [8 July 2022]

https://doi.org/10.1017/S0022215123000130 Published online by Cambridge University Press

13 Zhou Y, Liu C, Huang Y. Wireless power transfer for implanted medical application: a review *Energies* 2020;**13**:2837
14 Zeng FG, Rebscher S, Harrison W, Sun X, Feng H. Cochlear implants: system design, integration, and evaluation. *IEEE Rev Biomed Eng* 2008;**1**:115–42
15 BioMEMS Research group. In: https://flamenco.metu.edu.tr [8 July 2022]
16 Espressif Systems. In: https://docs.espressif.com/projects/espressif-esp-faq/en/latest/software-framework/ble-bt.html [11 July 2022]
17 Atmel. In: http://ww1.microchip.com/downloads/en/devicedoc/atmel-42426-smartconnect-samb11-soc_datasheet.pdf [11 July 2022]
18 White RL. Integrated circuits and multiple electrode arrays. In: M Merzenich, R Schindler, F Sooy, eds. *Proceedings of the First International Conference on Electrical Stimulation of the Acoustic Nerve as a Treatment for Profound Sensorineural Deafness in Man*. San Francisco: Velo-Bind, 1974;199–209
19 Avci P, Gupta A, Sadasivam M, Vecchio D, Pam Z, Pam N *et al.* Low-level laser (light) therapy (LLLT) in skin: stimulating, healing, restoring. *Semin Cutan Med Surg* 2013;**32**:41–52
20 Jin M-H, Wu W-J, Chen C-K, Chen Y-F, Wen C-M, Kao C-Y *et al.* Hierarchical sensor network architecture for stationary smart node supervision. *Proc of SPIE* 2004;53–63
21 Buti D, Pullano M, Papa E, Nygårds E, Ludvigsen L, Wadum J. Picasso's acrobat family in focus: an investigation of materials and techniques of an iconic work in the collection of the Gothenburg Museum of Art SN. *App Sci* 2020;**2**:1411
22 Cai D, Neyer A, Kuckuk R, Heise HM. Mid-infrared, near-infrared and ultraviolet–visible spectroscopy of PDMS silicone rubber for characterization of polymer optical waveguide materials. *J Mol Struct* 2010;**976**:274–81
23 Han D, Meng Z, Wu D, Zhang C, Zhu H. Thermal properties of carbon black aqueous nanofluids for solar absorption. *Nanoscale Res Lett* 2011;**6**:457
24 Kelley S, Rials T, Snell R, Groom L, Sluiter A. Use of near infrared spectroscopy to measure the chemical and mechanical properties of solid wood. *Wood Sci Tech* 2004;**38**:257–76
25 McLauchlin AR, Ghita O, Gahkani A. Quantification of PLA contamination in PET during injection moulding by in-line NIR spectroscopy. *Polym Test* 2014;**38**:46–52
26 Zidan HM, Abdelrazek EM, Abdelghany AM, Tarabiah AE. Characterization and some physical studies of PVA/PVP filled with MWCNTs. *J Mater Res Technol* 2019;**8**:904–13
27 Moghaddam L, Martin DJ, Halley PJ, Fredericks PM. Vibrational spectroscopic studies of laboratory scale polymer melt processing: application to a thermoplastic polyurethane nanocomposite. *Vibr Spectrosc* 2009;**51**:86–92
28 Prahl S. In: https://omlc.org/spectra/water/abs/index.html [6 August 2022]
29 Sarpeshkar R, Salthouse C, Sit JJ, Baker MW, Zhak SM, Lu TK *et al.* An ultra-low-power programmable analog bionic ear processor. *IEEE Trans Biomed Eng* 2005;**52**:711–27
30 Yigit HA, Ulusan H, Koc M, Yuksel MB, Chamanian S, Kulah H. Single supply PWM fully implantable cochlear implant interface circuit with active charge balancing. *IEEE Access* 2021;**9**:52642–53
31 Dejakum K, Piegger J, Plewka C, Gunkel A, Thumfart W, Kudaibergenova S *et al.* Medium-level laser in chronic tinnitus treatment. *Biomed Res Int* 2013;**2013**:324234

https://doi.org/10.1017/S0022215123000130 Published online by Cambridge University Press